

F° Freezer Temperature Log

Month/Year (Days 1-15) _____

Freezer Location/ID _____

VFC PIN _____

1

Record the Time and your Initials.

2

Check if Alarm triggered.

3

a. Record CURRENT, MIN, and MAX temperatures.*
b. Circle temperatures if TOO WARM. Refer to ranges.

4

Take action for one of the two ranges.*

Day of Month	Time	Initials
Example	8:00 a.m.	NN
	4:00 p.m.	NN
1	a.m.	
	p.m.	
2	a.m.	
	p.m.	
3	a.m.	
	p.m.	
4	a.m.	
	p.m.	
5	a.m.	
	p.m.	
6	a.m.	
	p.m.	
7	a.m.	
	p.m.	
8	a.m.	
	p.m.	
9	a.m.	
	p.m.	
10	a.m.	
	p.m.	
11	a.m.	
	p.m.	
12	a.m.	
	p.m.	
13	a.m.	
	p.m.	
14	a.m.	
	p.m.	
15	a.m.	
	p.m.	

Alarm	CURRENT	MIN	MAX
	-10.3	-20.2	-9.1
✓	2.4	-9.0	6.2

TOO WARM!

16° & warmer

15°

14°

13°

12°

11°

10°

9°

8°

7°

6°

5.1°

MAX TOO WARM?

- a. Press MEMORY CLEAR/ RESET buttons.
- b. Make sure freezer door is shut.
- c. Post "Do Not Use Vaccines" sign.
- d. Alert your supervisor.
- e. Report excursion at MyVFCvaccines.org.

OK

5.0°

4°

3°

2°

1°

0°

-1°

-2°

-3°

-4°

-5°

-6°

-7°

-8°

-9°

-10s°


-20s°

-30s°

-40 to -58.0°

MIN & MAX OK.

- a. Press MEMORY CLEAR/ RESET buttons.
- b. Make sure freezer door is shut. Done.



* If temperatures go below -58.0°F, report excursion at MyVFCvaccines.org.

* For data loggers, check instructions in product guide and at EZIZ.org/assets/docs/IMM-1206.pdf.

Supervisors Review log and sign when completed. Keep logs and excursion reports for 3 years.

I certify that temperatures recorded on this log are correct. All temperatures that were TOO WARM are circled and corrective action was taken. I understand that falsifying logs will result in vaccines being unusable—patients may need to be revaccinated and provider may need to replace vaccines.

On-site Supervisor's Name & Signature: _____ Date: _____

Staff Names with Initials: _____

F° Freezer Temperature Log

Month/Year (Days 16-31) _____

Freezer Location/ID _____

VFC PIN _____

1
Record the Time and your Initials.

2
Check if Alarm triggered.

3
a. Record CURRENT, MIN, and MAX temperatures.*
b. Circle temperatures if TOO WARM. Refer to ranges.

4
Take action for one of the two ranges.*

Day of Month	Time	Initials
16	a.m. / p.m.	-----
17	a.m. / p.m.	-----
18	a.m. / p.m.	-----
19	a.m. / p.m.	-----
20	a.m. / p.m.	-----
21	a.m. / p.m.	-----
22	a.m. / p.m.	-----
23	a.m. / p.m.	-----
24	a.m. / p.m.	-----
25	a.m. / p.m.	-----
26	a.m. / p.m.	-----
27	a.m. / p.m.	-----
28	a.m. / p.m.	-----
29	a.m. / p.m.	-----
30	a.m. / p.m.	-----
31	a.m. / p.m.	-----

Alarm	CURRENT	MIN	MAX
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----
-----	-----	-----	-----

TOO WARM!
16° & warmer

15°
14°
13°
12°
11°
10°
9°
8°
7°
6°
5.1°

MAX TOO WARM?

- Press MEMORY CLEAR/ RESET buttons.
- Make sure freezer door is shut.
- Post "Do Not Use Vaccines" sign.
- Alert your supervisor.
- Report excursion at MyVFCvaccines.org.

OK

5.0°
4°
3°
2°
1°
0°
-1°
-2°
-3°
-4°
-5°
-6°
-7°
-8°
-9°
-10s°
-20s°
-30s°
-40 to -58.0°

MIN & MAX OK.

- Press MEMORY CLEAR/ RESET buttons.
- Make sure freezer door is shut. Done.

If temperatures go below -58.0°F, report excursion at MyVFCvaccines.org.

* For data loggers, check instructions in product guide and at EZIZ.org/assets/docs/IMM-1206.pdf.

Supervisors Review log and sign when completed. Keep logs and excursion reports for 3 years.

I certify that temperatures recorded on this log are correct. All temperatures that were TOO WARM are circled and corrective action was taken. I understand that falsifying logs will result in vaccines being unusable—patients may need to be revaccinated and provider may need to replace vaccines.

On-site Supervisor's Name & Signature: _____ Date: _____

Staff Names with Initials: _____