Refrigerator Temperature Log

Month/Year (Days 1-15) _____

Refrigerator Location/ID _____

VFC PIN _____

1 Record Time and your Initials.

2 Check if Alarm triggered.

3 a. Record CURRENT, MIN, and MAX temperatures.* b. Circle if TOO WARM or TOO COLD. Refer to ranges.

4 Take action for one of the three ranges.*

Day of Month	Time	Initials	Alarm	CURRENT	MIN	MAX
Example	8:00 a.m.	NN		4.3	2.4	5.7
	4:00 p.m.	NN	✔	7.6	4.0	9.1
1	a.m. / p.m.					
2	a.m. / p.m.					
3	a.m. / p.m.					
4	a.m. / p.m.					
5	a.m. / p.m.					
6	a.m. / p.m.					
7	a.m. / p.m.					
8	a.m. / p.m.					
9	a.m. / p.m.					
10	a.m. / p.m.					
11	a.m. / p.m.					
12	a.m. / p.m.					
13	a.m. / p.m.					
14	a.m. / p.m.					
15	a.m. / p.m.					

TOO WARM!

16° & warmer
15°
14°
13°
12°
11°
10°
9°
8.1°

MAX TOO WARM?

a. Press MEMORY CLEAR/ RESET buttons.
b. Make sure refrigerator door is shut.
c. Post "Do Not Use Vaccines" sign.
d. Alert your supervisor.
e. **Report excursion at MyVFCvaccines.org.**

OK

8.0°
7°
6°
5°
4°
3°
2.0°

MIN & MAX OK.

a. Press MEMORY CLEAR/ RESET buttons.
b. Make sure refrigerator door is shut.

Done.

TOO COLD!

1.9°
1°
0°
-1°
-2°
-3°
-4°
-5°
-6° & colder

MIN TOO COLD?

a. Press MEMORY CLEAR/ RESET buttons.
b. Make sure refrigerator door is shut.
c. Post "Do Not Use Vaccines" sign.
d. Alert your supervisor.
e. **Report excursion at MyVFCvaccines.org.**

* For data loggers, check instructions in product guide and at EZIZ.org/assets/docs/IMM-1206.pdf.

Supervisors **Review log and sign when completed. Keep logs and excursion reports for 3 years.**

I certify that temperatures recorded on this log are correct. All temperatures that were TOO WARM or TOO COLD are circled and corrective action was taken. I understand that falsifying logs will result in vaccines being unusable—patients may need to be revaccinated and provider may need to replace vaccines.

On-site Supervisor's Name & Signature: _____ _____ Date: _____

Staff Names with Initials: _____ _____ _____

C° Refrigerator Temperature Log

Month/Year (Days 16-31) _____

Refrigerator Location/ID _____

VFC PIN _____

1 Record Time and your Initials.

2 Check if Alarm triggered.

3 a. Record CURRENT, MIN, and MAX temperatures.*
b. **Circle if TOO WARM or TOO COLD.** Refer to ranges.

4 Take action for one of the three ranges.*

Day of Month	Time	Initials	Alarm	CURRENT	MIN	MAX
16	a.m. / p.m.					
17	a.m. / p.m.					
18	a.m. / p.m.					
19	a.m. / p.m.					
20	a.m. / p.m.					
21	a.m. / p.m.					
22	a.m. / p.m.					
23	a.m. / p.m.					
24	a.m. / p.m.					
25	a.m. / p.m.					
26	a.m. / p.m.					
27	a.m. / p.m.					
28	a.m. / p.m.					
29	a.m. / p.m.					
30	a.m. / p.m.					
31	a.m. / p.m.					

TOO WARM!

16° & warmer
15°
14°
13°
12°
11°
10°
9°
8.1°

MAX TOO WARM?

a. Press MEMORY CLEAR/ RESET buttons.
b. Make sure refrigerator door is shut.
c. Post "Do Not Use Vaccines" sign.
d. Alert your supervisor.
e. **Report excursion at MyVFCvaccines.org.**

OK

8.0°
7°
6°
5°
4°
3°
2.0°

MIN & MAX OK.

a. Press MEMORY CLEAR/ RESET buttons.
b. Make sure refrigerator door is shut.

Done.

TOO COLD!

1.9°
1°
0°
-1°
-2°
-3°
-4°
-5°
-6° & colder

MIN TOO COLD?

a. Press MEMORY CLEAR/ RESET buttons.
b. Make sure refrigerator door is shut.
c. Post "Do Not Use Vaccines" sign.
d. Alert your supervisor.
e. **Report excursion at MyVFCvaccines.org.**

* For data loggers, check instructions in product guide and at EZIZ.org/assets/docs/IMM-1206.pdf.

Supervisors **Review log and sign when completed. Keep logs and excursion reports for 3 years.**

I certify that temperatures recorded on this log are correct. All temperatures that were TOO WARM or TOO COLD are circled and corrective action was taken. I understand that falsifying logs will result in vaccines being unusable—patients may need to be revaccinated and provider may need to replace vaccines.

On-site Supervisor's Name & Signature: _____ _____ Date: _____

Staff Names with Initials: _____ _____ _____