

COVID-19 Temperature Log

MONTH & YEAR STORAGE UNIT LOCATION/ID PROVIDER PIN

--	--	--

Day	Time	Initials	Alarm	Current	Min	Max	Incident IDs
Exam- ple	8:00 am	NN		40.5	38.1	43.7	
	4:00 pm	NN	✓	37.4	33.0	39.2	12345
1	am						
	pm						
2	am						
	pm						
3	am						
	pm						
4	am						
	pm						
5	am						
	pm						
6	am						
	pm						
7	am						
	pm						
8	am						
	pm						
9	am						
	pm						
10	am						
	pm						
11	am						
	pm						
12	am						
	pm						
13	am						
	pm						
14	am						
	pm						
15	am						
	pm						

Identify product and storage unit:

All Pfizer-BioNTech:

- ULT Freezer: until expiration -130° to -76°F (-90° to -60°C)
- Refrigerator: 10 weeks 36° to 46°F (2° to 8°C)

All Moderna:

- Freezer: until expiration -58° to 5°F (-50° to -15°C)
- Refrigerator: 30 days 36° to 46°F (2° to 8°C)

Novavax:

- Refrigerator: until expiration 36° to 46°F (2° to 8°C)

Check temperatures twice a day:

1. Fill out clinic details in header.
2. Record the time and your initials.
3. Record a check if alarm went off.
4. Record Current, MIN, and MAX.
5. Clear MIN/MAX.
6. Ensure data logger is recording.

IF ALARM WENT OFF:

1. Clear MIN/MAX and alarm symbol.
2. Post "Do Not Use Vaccines" sign.
3. Alert your supervisor.
4. Report excursion to myCAvax.
5. Record myCAvax Batch Excursion # under Incident IDs.
6. Ensure data logger is recording.



Keep all temperature logs and data files for three years.

BAP: (833) 502-1245

Staff Names and Initials:

On-Site Supervisor's Review – When log is complete, check all that apply:

- Temperatures were recorded twice daily.
- I reviewed data files to find any missed excursions. Download date: _____
- Any excursions were reported.

Name: _____ Signature: _____ Date: _____

COVID-19 Temperature Log



MONTH & YEAR STORAGE UNIT LOCATION/ID PROVIDER PIN

--	--	--

Day	Time	Initials	Alarm	Current	Min	Max	Incident IDs
16	am						
	pm						
17	am						
	pm						
18	am						
	pm						
19	am						
	pm						
20	am						
	pm						
21	am						
	pm						
22	am						
	pm						
23	am						
	pm						
24	am						
	pm						
25	am						
	pm						
26	am						
	pm						
27	am						
	pm						
28	am						
	pm						
29	am						
	pm						
30	am						
	pm						
31	am						
	pm						

Identify product and storage unit:

All Pfizer-BioNTech:

- ULT Freezer: until expiration
-130° to -76°F (-90° to -60°C)
- Refrigerator: 10 weeks
36° to 46°F (2° to 8°C)

All Moderna:

- Freezer: until expiration
-58° to 5°F (-50° to -15°C)
- Refrigerator: 30 days
36° to 46°F (2° to 8°C)

Novavax:

- Refrigerator: until expiration
36° to 46°F (2° to 8°C)

Check temperatures twice a day:

1. Fill out clinic details in header.
2. Record the time and your initials.
3. Record a check if alarm went off.
4. Record Current, MIN, and MAX.
5. Clear MIN/MAX.
6. Ensure data logger is recording.

IF ALARM WENT OFF:

1. Clear MIN/MAX and alarm symbol.
2. Post "Do Not Use Vaccines" sign.
3. Alert your supervisor.
4. Report excursion to myCAvax.
5. Record myCAvax Batch
Excursion # under Incident IDs.
6. Ensure data logger is recording.



Keep all temperature logs and data files for three years.

BAP: (833) 502-1245

Staff Names and Initials:

On-Site Supervisor's Review – When log is complete, check all that apply:

- Temperatures were recorded twice daily.
- I reviewed data files to find any missed excursions. Download date: _____
- Any excursions were reported.

Name: _____ Signature: _____ Date: _____